



TEST REPORT TUCHENG

ORIGINAL

Date of Issue: Nov.17,2020 Date of Receipt and Test Start: Oct.28,2020

Report No.: TFF9J683 Quantity: 1PC Page Order/Pages: (P1/2) Ref. No.: NIL

Report Title: ATARAXIS MEDICAL SUPPLY CO., LTD.(U3494) Item: Mask

Address: 2F., No. 14, Ln. 122, Sec. 2, Xiwan Rd., Xizhi Dist., New Taipei City 221, Taiwan (R.O.C.)

Test Items		Test Results	Test Methods
*Synthetic Blood Penetration Pressure:80 mmHg	1	None Seen	CNS 14774 T5017-2018 9.1
	2	None Seen	CNS 14776 T4038-2003
	3	None Seen	
	4	None Seen	
	5	None Seen	
	6	None Seen	
	7	None Seen	
	8	None Seen	
	9	None Seen	
	10	None Seen	
	11	None Seen	
	12	None Seen	
	13	None Seen	

Note: Sample description is given by the client: ATARAXIS 3D DISPOSABLE FACE MASK

Note: The test report is the test results issued by the testing institution as requested by the consignor, it shall not determine the legitimacy of the product.

Note: "*" indicates that the tests are accredited by the Ministry of Health and Welfare in Taiwan, and are conducted according to the accredited test methods.

- Note: 1.This report is only responsible for the submitted sample(s), which will be kept for one month period.
- 2.This report cannot be reproduced in any way, except in full context, without the prior approval in writing of this Department of Testing and Certification.
- 3.The test report should not be used for public advertisement and commercial promotion.

Authorized by president of Jui-hung kao
Taiwan Textile Research Institute

Director,
Department of Testing and
Certification

Department of Testing and Certification, Taiwan Textile Research Institute
No.6, Chengtian Rd., Tucheng Dist., New Taipei City 23674, Taiwan (R.O.C.)
Tel : +886-2-22670321 ext. 7107, 7110
Fax : +886-2-22675108 , +886-2-22689839



TEST REPORT TUCHENG

ORIGINAL

Date of Issue: Nov. 17, 2020 Date of Receipt and Test Start: Oct. 28, 2020

Report No.: TFF9J683 Quantity: 1PC Page Order/Pages: (P2/2) Ref. No.: NIL

Report Title: ATARAXIS MEDICAL SUPPLY CO., LTD. (U3494) Item: Mask

Address: 2F., No. 14, Ln. 122, Sec. 2, Xiwan Rd., Xizhi Dist., New Taipei City 221, Taiwan (R.O.C.)



- Note: 1.This report is only responsible for the submitted sample(s), which will be kept for one month period.
2.This report cannot be reproduced in any way, except in full context, without the prior approval in writing of this Department of Testing and Certification.
3.The test report should not be used for public advertisement and commercial promotion.

Authorized by president of
Taiwan Textile Research Institute

Jui-hung Kao

Director,
Department of Testing and Certification
Department of Testing and Certification
No.6, Chengtian Rd., Tucheng Dist., New Taipei City 23674, Taiwan (R.O.C.)
Tel : +886-2-22670321 ext. 7107, 7110
Fax : +886-2-22675108 , +886-2-22689839